

### Sheffield – testing – head (art. no. 1691)

to perform various breathing resistance tests, e.g.


- according to DIN EN 149, §§ 8.9.2, 8.9.3
- according to DIN EN 13274, §§ 7.3.2, 7.3.3

with concentric aspiration-, pressure- and measuring tube in the mouth area

90° swingable:

to the front, to the back, to the left, to the right

with hose connections nominal diameter 15 for inhalation and exhalation and nominal diameter 4 for pressure measurement



Further information at

[www.drwiesner.de](http://www.drwiesner.de)

**Wiesner Prüftechnik GmbH**

Weststrasse 4  
DE-73630 Remshalden – Germany

Phone: +49 7151 9736 - 0  
E-Mail: [info@drwiesner.de](mailto:info@drwiesner.de)